

INTRODUCTION

The *Performance and Responsibility Guidelines for the Nuclear Medicine Technologist* were developed initially by the Socioeconomic Affairs Committee and approved in 1994. Over this past year, the Academic Affairs and Socioeconomic Affairs Committees have revised the 1994 guidelines to bring them in line with current practice for nuclear medicine technologists. The SNM-TS National Council and House of Delegates approved this latest version at their recent meeting in Las Vegas.

Socioeconomic Affairs Committee:

Denise Merlino (Chair), Becky Cacciatore, Chris Carlson, Mickey Clarke, Valerie Cronin, James Fletcher, MD, Kathleen Krisak, Ken McKusick, MD, Ann Orenak, April Russell, Lynne Roy, Mary Jo Struttman, Sharon Surrel, Howard Teng, LisaAnn Trembath, Kristen Waterstram-Rich, Cynthia Wharton

Academic Affairs Committee: Kristen Waterstram-Rich (Chair), Ken Blasius, Paul Christian, Patti Corrigan, Elpida Crawford, Elaine Cuklanz, James Greene, Joni Herbst, Susan Laffin, Brenda McSherry, Miriam Miller, Wanda Mundy, Kathleen Murphy, Lorie Overbey, Carole South-Winter, Sharon Surrel, Susan Weiss, Patricia Wells

THE GUIDELINES

The spectrum of nuclear medicine technology skills and responsibilities varies widely across the country. The broad descriptions of this document will provide a basis for determining the areas of knowledge and of performance for the nuclear medicine technologist. The documents used in the revision and development of these guidelines were the *Performance and Responsibility Standards for the Nuclear Medicine Technologist* (1994), *Nuclear Medicine Technology Certification Board Report: Equipment and Procedures in Current Practice*, *Nuclear Medicine Technology Certification Board Critical Task Analysis Report* and the *Essentials and Guidelines for an Accredited Educational Program for the Nuclear Medicine Technologist* (1997). These guidelines should be considered a helpful checklist of those skills necessary to perform a variety of nuclear medicine procedures. While the editors tried to be complete, nuclear medicine technology is a dynamic and evolving field; therefore, any list is likely to be partially obsolete as soon as it is issued. In addition, this document is not designed to be a "how to" description for any of the listed activities, nor is it intended to modify or alter existing tort law.

Nuclear medicine is the field of medicine that uses unsealed radioactive materials in the diagnosis and treatment of disease. This includes the administration of a radiopharmaceutical to a patient for the therapeutic treatment and/or the imaging of the radiopharmaceutical distribution in an organ or area of interest within the patient.

Nuclear Medicine Technology

The practice of nuclear medicine technology encompasses multidisciplinary skills. The responsibilities of the nuclear medicine technologist include an empathetic and instructional approach to patient care, the preparation, calibration and administration of radiopharmaceuticals and pharmaceuticals, (under the direction of an authorized user) the performance of quality control procedures and the operation of imaging, laboratory and computer instrumentation.

Continuing education is a necessary component in maintaining the skills required to perform the duties and tasks of the nuclear medicine technologist.

In order to perform these tasks, the nuclear medicine technologist

Performance and Responsibility Guidelines for the Nuclear Medicine Technologist: 1998 Revision

must successfully complete didactic and clinical training. Recommended course work includes but is not limited to: anatomy, physiology, pathophysiology, chemistry, physics, mathematics, computer applications, biomedical sciences, ethics, and radiation science health and safety. Direct patient contact hours are obtained by training in a clinical setting.

Formal education programs in nuclear medicine technology are accredited by the Joint Review Committee on Educational Programs in Nuclear Medicine Technology (JRCNMT). Graduates of accredited pro-

grams are eligible to take the certification examination offered by the Nuclear Medicine Technology Certification Board (NMTCB) and/or American Registry of Radiologic Technologists (ARRT).

The scope of performance in nuclear medicine technology includes, but is not limited to the following areas and responsibilities:

Patient Care: Requires the exercise of judgment to assess and respond to patient's needs prior to, during and after visits in the nuclear medicine department.

Quality Control: Requires the evaluation and maintenance of a quality control program for instrumentation to insure its credibility and reliability.

Diagnostic Procedures: Requires the utilization of appropriate technique to insure quality diagnostic images and/or laboratory results.

Radiopharmaceuticals: Involves the procurement, preparation, quality control, calculation, identification, documentation, administration, disposal, storage, and safe handling of such materials.

Radionuclide Therapy: Assists an authorized user in the application and management of a therapeutic radionuclide treatment.

Radiation Safety: Educates the public and uses techniques that will minimize radiation exposure to the patient, general public, and health care personnel consistent with the ALARA (as low as is reasonably achievable) concept.

I. Patient Care

- A. A nuclear medicine technologist provides patient care:
 1. By providing for proper comfort and care to the patient prior to, during and after a procedure, including the monitoring of IVs, oxygen supplies, and drains. In addition, monitors patients who are under conscious sedation, in those facilities who approve and document competency of all monitoring staff. [In accordance with the American Society of Anesthesiology's (ASA) guidelines for conscious sedation.]
 2. By establishing and maintaining good communication with patients (i.e., introducing themselves, explaining the procedures, answering questions).
 3. By behaving in a professional manner in consideration of patients' rights and conducting himself/herself at all times in a manner not to bring harm to the patient.
 4. By providing functionally safe and sanitary conditions for the patient in compliance with universal protection policies.
 5. By recognizing and responding to an emergency condition by:
 - a. Initiating a call for assistance.
 - b. Monitoring and recording physiologic data (i.e., ECG, pulse rate, respiratory rate).
 - c. Administering cardiopulmonary resuscitation when necessary.
 - d. Maintaining intravenous fluids, oxygen and other life-support assistance until an emergency code team arrives.

- B. A nuclear medicine technologist prepares the patient:
1. By verifying patient identification, last menstrual period and written orders for the procedure.
 2. By checking for contraindications and obtaining a pertinent history.
 3. By assuring any pre-study preparation has been completed, i.e., hydration, voiding, bowel cleansing, suspension of interfering drugs.
 4. By assuring informed consent has been obtained, when necessary.
 5. By explaining the procedure to the patient.
 6. By checking patient clothing and linen for objects that may cause artifacts in the images.
 7. By waiting an appropriate length of time after the administration of a radiopharmaceutical to begin the procedure.
 8. By checking the indication for the study for appropriateness, and following up with the authorized user and/or referring physician, to assure that the proper study is being performed.
- C. A nuclear medicine technologists performs administrative procedures:
1. By maintaining an adequate volume of medical/surgical supplies, radiopharmaceuticals and film to ensure that a patient procedure may be performed whenever necessary.
 2. By scheduling patient procedures with timely arrangements.
 3. By determining the appropriate sequence for multiple procedure requests.
 4. By maintaining appropriate records of patient doses, quality control procedures, patient reports, and other required records.
 5. By revising and developing recordable event guidelines in collaboration with an authorized user ("events" replaced the word "procedures" to better follow NRC terminology).
 6. By participating in the quality assurance program.
- II. Nuclear Instrumentation—Quality Control**
- A. A nuclear medicine technologist evaluates the performance of scintillation cameras:
1. By obtaining uniformity images:
 - a. Selecting a radionuclide source of appropriate type, size, (if necessary), quantity and energy.
 - b. Selecting an appropriate pulse height analyzer (PHA) photpeak and window.
 - c. Obtaining uniformity images using standardized imaging parameters.
 - d. Evaluating the images qualitatively and/or quantitatively in comparison to the manufacturer's specifications and the performance requirements based on the studies for which unit is used.
 - e. Identifying the source of any nonuniformity (i.e., checking collimator, PHA peak setting).
 - f. Initiating corrective action when necessary.
 - g. Maintaining required records for the quality control program
 2. By performing a detector linearity evaluation:
 - a. Selecting a radionuclide, a linearity phantom and obtaining images.
 - b. Identifying any nonlinear distortion in the image
 - c. Determining the source of nonlinearity (i.e., detector-source geometry).
 - d. Initiating corrective action when necessary.
 3. By performing spatial resolution checks:
 - a. Selecting an appropriate radionuclide.
 - b. Choosing a phantom that is compatible with the specified resolution of the camera.
 - c. Analyzing the resulting images for degradation of resolution.
 - d. Initiating corrective action when necessary.
 - e. Maintaining required records for the quality control program.
 4. By conducting sensitivity checks:
 - a. Selecting a source with an appropriate level of activity and half-life.
 - b. Assuring identical geometry, source placement and measurement parameters for repetitive checks.
 - c. Evaluating results.
 - d. Initiating corrective action when necessary.
 - e. Maintaining required records for the quality control program.
 5. By performing SPECT quality control procedures:
 - a. Obtaining a high-count uniformity flood.
 - b. Obtaining a center-of-rotation correction.
 - c. Verifying energy correction and spatial coordinates.
 - d. Verifying multi-head detector alignment.
 - e. Evaluating reconstruction results of phantom acquisition.
 - f. Analyzing the results for degradation.
 - g. Initiating corrective action when necessary.
 - h. Maintaining required records for the quality control program.
 6. By checking computer parameter settings and data interface:
 - a. Assuring camera and computer register same count rate at max frame rate.
 - b. Verifying the camera and computer have the same image orientation.
 - c. Obtaining a dead-time measurement on the computer.
 - d. Verifying accuracy of ECG gating.
 - e. Performing pixel calibration.
 7. By checking the analog and/or digital recording device(s):
 - a. Performing a lens focus check (i.e., CRT).
 - b. Checking and adjusting imaging device for contrast and brightness (i.e., densitometry).
 - c. Assessing integrity of imaging device.
 - d. Maintaining cleanliness of all equipment (i.e., lens, fan covers).
 8. A nuclear medicine technologists actively participates in total quality management/continuous quality improvement programs:
 - a. By identifying procedures to be analyzed.
 - b. By gathering and presenting data in appropriate formats.
 - c. By analyzing data and recommending changes.
- B. A nuclear medicine technologist evaluates the performance of NaI (TI) scintillation probes and well counters:
1. By calibrating a spectrometer with a calibrated long half-life radionuclide source.
 2. By determining energy resolution.
 3. By conducting sensitivity measurements at appropriate energies.
 4. By checking background and determining the cause for levels greater than established normal levels.

5. By conducting a chi-square test.
6. By maintaining required records for quality control programs.
- C. A nuclear medicine technologist operates survey meters:
 1. By ensuring calibration is completed with an approved agent.
 2. By performing a reference check-source test and comparing with previous results.
 3. By maintaining required records for quality control program.
- D. A nuclear medicine technologist evaluates the operation of a dose calibrator:
 1. By verifying the calibration.
 2. By ascertaining linearity over the entire range of radionuclide activity to be measured and determining correction factors when necessary.
 3. By testing for significant geometric variation in activity measured as a function of sample volume or configuration and determining correction factors when necessary.
 4. By determining accuracy.
 5. By determining precision (constancy).
 6. By maintaining required records for the quality control program.
- E. A nuclear medicine technologist operates and maintains film processors:
 1. By monitoring and recording sensitometry and temperature of water and dryer daily.
 2. By maintaining required records for quality control program.
 3. By performing data collection, processing and analysis:
 - a. Reviewing images to assure that correct information is supplied.
 4. By assisting the physician in cardiac stress testing when performed in conjunction with nuclear medicine procedures:
 - a. Preparing patient's skin and placing ECG leads appropriately.
 - b. Recognizing and being responsive to any changes that may occur on either a resting or stress ECG.
 - c. Recognizing the parameters that should terminate a cardiac stress study.
 - d. Recognizing ECG patterns that are appropriate for gating.
 5. By performing data collection, processing and analysis:
 - a. Performing data collection, processing and analysis in accordance with established protocols.
 - b. Exercising independent judgment in selecting appropriate images for processing.
 - c. Selecting appropriate filters, frequency cutoff and attenuation correction when reconstructing SPECT images.
 - d. Defining regions of interest (ROIs) with reproducible results and correctly applying background subtraction.
 - e. Performing computer data manipulations as required by standard nuclear medicine procedures, i.e., activity curve generation, quantitation, SPECT slice production.
 - f. Labeling processed images to reflect anatomical positioning, ROIs, etc.
 - g. Preserving and retrieving data from storage media.

III. Diagnostic Procedures

- A. A nuclear medicine technologist performs imaging procedures:
 1. By selecting imaging parameters:
 - a. Selecting and preparing the instrument for the procedure.
 - b. Selecting appropriate parameters for digital and/or analog acquisition.
 - c. Recognizing artifacts that are due to instrumentation malfunction and initiating appropriate action.
 2. By administering radiopharmaceuticals and/or pharmaceuticals using universal precaution techniques:
 - a. Verifying patient identity prior to the administration of medication or radiopharmaceuticals.
 - b. Determining route of administration according to established protocol (i.e., subcutaneous, intramuscular, intravenous, etc.).
 - c. Establishing and/or verifying venipuncture access using aseptic technique.
 - d. Using and maintaining established venous access routes (i.e., heparin infusion, IMED).
 - e. Establishing patient patterned breathing when introducing radiopharmaceuticals (i.e., inhalants or aerosols).
 - f. Administering oral radiopharmaceuticals.
 - g. Documenting medications and/or radiopharmaceutical administrations on a patient's permanent record.
 3. By positioning the patient and obtaining images:
 - a. Performing imaging views according to established protocols and acquiring additional views to optimize information content.
 - b. Placing the patient in correct position using supportive materials and immobilizer as necessary.
 - c. Exercising independent judgment in positioning a patient or detector unit to best demonstrate pathology and to adapt to the patient's limitations.
 - d. Indicating appropriate anatomic landmarks for each view of the procedure.
 - e. Reviewing images to assure that correct information is supplied.
- B. A nuclear medicine technologist performs nonimaging in vivo and/or radioassay studies:
 1. By operating laboratory equipment:
 - a. Checking accuracy, precision and operation of pipetting device.
 - b. Using microhematocrit centrifuge and determining hematocrit.
 2. By preparing doses and guidelines:
 - a. Quantitating dose by:
 1. Determining decay factor and calculating remaining activity.
 2. Determining volume necessary to deliver activity for the prescribed dose.
 3. Drawing dose into syringe using appropriate techniques and materials.
 4. Dispensing appropriate quantity of liquid or capsules, as necessary, for the prescribed dose.
 5. Confirming calculated activity by using a dose calibrator.
 - b. Preparing standard by:
 1. Choosing appropriate volumetric or gravimetric techniques to dilute standard.
 2. Adding radioactive material identical to that given the patient q.s. (quantify sufficient) to appropriate volume.
 3. Diluting capsule in appropriate solvent, if necessary, for preparing a standard.
 3. By collecting proper specimen for procedures using universal precaution techniques:
 - a. Collecting blood samples by:
 1. Selecting proper supplies (i.e., needles, syringes, evacuated tubes, anticoagulants, etc.).

2. Labeling patient demographics on collection containers.
 3. Performing venipuncture at appropriate time intervals using aseptic technique.
 4. Adding hemolyzing compounds or anticoagulants to samples when necessary.
 5. Centrifuging blood and separating blood components, as required.
 6. Storing aliquot of serum, plasma or whole blood according to protocol.
- b. Collecting urine samples by:
1. Instructing patient and nursing staff as to correct method and time of urine collection.
 2. Aliquoting urine sample and measuring total urine volume.
 3. Measuring specific gravity of urine, if required.
 4. Recognizing and documenting all technical circumstances which would produce invalid results.
4. By performing calculations:
- a. Subtracting room or patient background from appropriate samples.
 - b. Applying appropriate formulas, including conversion and dilution factors.
 - c. Calculating results according to procedure used.
 - d. Plotting graph, if necessary, and determining half time by extrapolating to zero time.
 - e. Reporting both patient calculated values and normal range of specific procedures used.
 - f. Evaluating results for potential error.
5. By managing biohazardous waste using disposal methods adopted as facility policy.

IV. Radiopharmaceuticals

- A. A nuclear medicine technologist obtains and maintains radiopharmaceutical products and adjunct supplies:
1. By anticipating and procuring a sufficient supply of radioactive drugs for an appropriate time period in accordance with anticipated need and license possession limits.
 2. By storing drugs and supplies in a manner consistent with labeled product safeguards and with radiation safety considerations.
 3. By performing and documenting radiation wipe tests upon receipt of radioactive materials.
 4. By recording receipt of radioactive materials in a permanent record.
 5. By following DOT and radiation safety guidelines in the transport, receipt and shipment of radioactivity.
- B. A nuclear medicine technologist prepares and verifies quality of radiopharmaceuticals under the direction of an authorized user:
1. By employing aseptic technique for manipulation of injectable products.
 2. By assembling and maintaining radionuclide generators.
 3. By eluting radionuclide generators according to manufacturer's specification.
 4. By verifying radionuclide purity of generator eluates.
 5. By selecting and preparing radiopharmaceuticals in accordance with manufacturer's specifications.
 6. By measuring and calculating activity of the radionuclide with a dose calibrator.
 7. By confirming the quality of a radiopharmaceutical in accordance

with accepted techniques and official guidelines, i.e. radiochemical purity, physical appearance.

8. By preparing labeled blood cells (i.e., ^{111}In WBC) in accordance with established protocols.
 9. Recording use and/or disposition of all radioactive materials in a permanent record.
- C. nuclear medicine technologist is responsible for the identification and labeling of all radiopharmaceutical preparations:
1. By labeling the container as required by regulation.
 2. By recording radiopharmaceutical and medication information on a patient's administration form and permanent preparation records.
 3. By labeling and segregating radioactive waste and recording this information in a permanent record.
- D. A nuclear medicine technologist prepares individual dosages under the direction of an authorized user:
1. By applying radioactive decay calculations to determine required volume or unit form necessary to deliver the prescribed radioactive dose.
 2. By selecting and preparing prescribed dosages and entering this information on a patient's administration form and other permanent records.
 3. By labeling the dose for administration.
 4. By checking the dose activity prior to administration in a dose calibrator and comparing this measurement against the identification label of the dose's immediate container.

V. Radionuclide Therapy

- A. Nuclear medicine technologist assists an authorized user in the preparation and applications of therapeutic radionuclides:
1. assuring the correct radiopharmaceutical and dosage is prepared.
 2. By following the NRC mandated quality management program in effect at the facility in regard to patient identification and the use of therapeutic radionuclides.
 3. By preparing and/or coordinating environmental preparations (i.e., decontamination supplies).
 4. By observing prescribed radiation safety procedures during the preparation and the administration of such treatment.
 5. By assisting the authorized user in supplying proper patient care instructions to hospital staff, patient, and/or caregivers.
 6. By conducting and documenting radiation surveys of designated patient areas, when indicated.
 7. By supplying hospital staff, patient, and/or caregivers with proper instructions on handling and disposal of all contaminated supplies, when necessary.

VI. Radiation Safety

- A. A nuclear medicine technologist under supervision of an authorized user and/or under the supervision of, or serving as the radiation safety officer maintains compliance with local, state or federal regulations in radiation safety practices:
1. By notifying appropriate authority when changes occur in the radiation safety program.
 2. By assisting in the preparation of license amendments, when necessary.
 3. By keeping up to date on regulatory changes and by complying with all applicable regulations.
 4. By maintaining required records.
 5. By posting appropriate signs in designated areas.

SNM-TS SPECIAL REPORT

6. By following regulations regarding receipt and disposition of all radionuclides.
 7. By carrying out a program to follow regulations regarding therapeutic doses and follow-up.
 8. By recommending purchase of protection equipment to meet regulations.
 9. By packaging radioactive material according to regulations and keeping accurate records of transfer.
- B. A nuclear medicine technologist follows appropriate protection procedures:
1. By using personnel monitoring devices (dosimeters, film badges, TLDs, etc.):
 - a. Reviewing monthly personnel exposure records in regard to maximum permissible dose limits.
 - b. Taking appropriate measures to reduce exposure, when necessary.
 - c. Notifying proper authorities of excessive exposure upon occurrence.
 2. By selecting and using proper shielding to reduce radiation exposure.
 3. By using proper methods for storage and disposal of radioactive materials.
 4. By identifying and using proper procedures for those radionuclides that pose special hazards (i.e., ^{90}Sr , ^{131}I).
 5. By performing a bioassay on a technologist as per state and/or federal regulations.
 6. By working in a safe, but timely manner in order to decrease radiation exposure.
 7. By following ALARA principle.
 8. By reviewing personal monitoring device readings to determine if radiation exposure can be further reduced.
 9. By working in a manner that minimizes potential contamination of patients, technologists, the public and work areas.
- C. A nuclear medicine technologist performs surveys:
1. By ensuring that instruments are calibrated at regular intervals, or after repairing according to regulations.
 2. By setting frequency and locations for surveys and following schedules.
 3. By using appropriate survey meters for each type and level of activity.
 4. By following regulations regarding personnel surveys and reporting to the designated physician or radiation safety officer.
 5. By performing constancy checks on survey meters.
 6. By performing wipe tests where applicable.
 7. By performing leak tests on sealed sources, when so authorized.
 8. By recording data in standard format.
 9. By evaluating results of wipe tests and area surveys to determine if action is required.
 10. By notifying the radiation safety officer when actions are required.
- D. A nuclear medicine technologist performs decontamination procedures:
1. By wearing appropriate clothing and foot covering as necessary.
 2. By blocking access to area and confining a spill.
 3. By removing contamination or reducing the activity to acceptable levels.
 4. By monitoring the area and personnel and repeating decontamination procedure until activity levels are acceptable.
 5. By closing off all areas of fixed contamination that are above acceptable levels.
 6. By identifying, storing or disposing of contaminated material in accordance with regulations.
 7. By maintaining adequate records concerning cleanup.
 8. By notifying appropriate authority (i.e., radiation safety officer) in the event of possible overexposure or other violations of regulations.
- E. A nuclear medicine technologist disposes of radioactive waste:
1. By maintaining appropriate records.
 2. By disposing of waste properly according to license specifications.
 3. By maintaining long- and short-term storage areas according to regulation.
- F. A nuclear medicine technologist participates in a hospital's in-service program to instruct other personnel about radiation hazards and principles of radiation safety:
1. By teaching concepts:
 - a. Types of ionizing radiation.
 - b. The biological effects of ionizing radiation.
 - c. Limits of dose, exposure and radiation effect.
 - d. Concepts of low-level radiation and health.
 - e. Concept of risk versus benefit.
 2. By providing instruction on appropriate radiation safety measures.
 3. By providing instruction on proper emergency procedures to be followed until radiation safety personnel arrive at the site of accident or spill.

REFERENCES

- Nuclear Medicine Technology Certification Board report: equipment and procedures in current practice.* Atlanta: Nuclear Medicine Technology Certification Board; Sep. 1993.
- Blondeau K, Hartnett S, Pickett M, Bridges J. *Nuclear Medicine Technology Certification Board critical task analysis report.* Atlanta: Nuclear Medicine Technology Certification Board; 1991.
- Essentials and guidelines for an accredited educational program for the nuclear medicine technologist,* revised. Salt Lake City, UT: Joint Review Committee on Educational Programs in Nuclear Medicine Technology; 1991.
- Socioeconomic Affairs Committee, Matthews J, Bonds S, Cianci M, Cooper J, Hanis K, Hodges H, Kraft R, Langan J, Porter J, Price D, Wald D. *Performance and responsibility: guidelines for the nuclear medicine technologist.* New York: Society of Nuclear Medicine-Technologist Section; Sept. 1982.
- Socioeconomic Affairs Committee, Higgins D, Hall A, Clarke M, Mayle L, Montanaro-Sebel A. *Performance and responsibility: guidelines for the nuclear medicine technologist.* Reston, VA: Society of Nuclear Medicine-Technologist Section; Sept. 1994.